

# 计量标准考核证书

Certificate for Examination of Measurement Standard

[ 2018 ] 长 量 标 企 证 字 第 247 号

根据《中华人民共和国计量法》，按照《计量标准考核规范》的要求，考核合格，特发此证。

This is to certify that the measurement standard conforms with the requirements of the "Rule for the Examination of Measurement Standards" according to "the Law on Metrology of the People's Republic of China".

建标单位名称 长沙中通检测技术服务有限公司  
Possessor of the Measurement Standard

计量标准名称 天平检定装置  
Name of Measurement Standard

代码 12114600  
Code

测量范围 1mg~5000g  
Measuring Range

不确定度或准确度等级 F1等级  
或最大允许误差  
Uncertainty or Accuracy Class or  
Maximum Permissible Error

保存地点 湖南省长沙市岳麓区学士街道茯苓路30号六楼  
Installed in

发证机关 (印章)  
The Issuing Authority

发证日期 2018年 06月 15日  
Date of Issue

有效期至 2022年 06月 14日  
Date of Expiry

名称 Name	型号 Type	测量范围 Measuring Range		不确定度或准确度等级 或最大允许误差 Uncertainty or Accuracy Class or Maximum Permissible Error	制造厂及 出厂编号 Manufacturer and Series Number
		无磁砝码	1mg-500g	1mg-500g	F1等级
无磁砝码	1kg、2kg、5kg	1kg、2kg、5kg	F1等级	蓬莱鸿鑫平衡器 有限公司 4143	
温湿度表	HTC-1	温度：(0~50)℃ 湿度：(0~90)%RH	温度： $U=0.5^{\circ}\text{C}$ $k=2$ 湿度： $U=3\%RH$ $k=2$	/	
标准水银温度计	(0~50)℃, 0.1℃ 分度	(0~50)℃	$U=0.05^{\circ}\text{C}$ $k=2$	/2-40	
名称 Name	测量范围 Measuring Range		不确定度或准确度等级 或最大允许误差 Uncertainty or Accuracy Class or Maximum Permissible Error	依据的计量检定规程 或计量技术规范的编号及名称 Verification Regulation or Measurement Technical Specification and Its Code	
	天平	1mg~5000g	II 级及以下	JJG1036-2008 《电子天平检定规程》	

计量  
标准器

Measurement  
Standard

主要  
配套  
设备

Main  
Auxiliary  
Equipment

开展的  
检定  
或校  
准项  
目

Verification  
or  
Calibration  
Items