

计量标准考核证书

Certificate for Examination of Measurement Standard

[2019] 长 量 标 企 证 字 第 278 号

根据《中华人民共和国计量法》，按照《计量标准考核规范》的要求，考核合格，特发此证。

This is to certify that the measurement standard conforms with the requirements of the "Rule for the Examination of Measurement Standards" according to "the Law on Metrology of the People's Republic of China".

建标单位名称 长沙中通检测技术服务有限公司
Possessor of the Measurement Standard


计量标准名称 温度二次仪表检定装置
Name of Measurement Standard

代码 04117100
Code

测量范围 (-200~1300) °C
Measuring Range

不确定度或准确度等级 0.02级
或最大允许误差
Uncertainty or Accuracy Class or
Maximum Permissible Error

保存地点 湖南省长沙市岳麓区学士街道茯苓路30号六楼本公司实验室
Installed in



发证日期 2019 年 03 月 15 日
Date of Issue

有效期至 2023 年 03 月 14 日
Date of Expiry

	名 称 Name	型 号 Type	测 量 范 围 Measuring Range	不 确 定 度 或 准 确 度 等 级 或 最 大 允 许 误 差 Uncertainty or Accuracy Class or Maximum Permissible Error	制 造 厂 及 出 厂 编 号 Manufacturer and Series Number
	计 量 标 准 器 Measurement Standard	特稳携式校准仪	JY822	(-200~1300) °C	0.02级
主 要 配 套 设 备 Main Auxiliary Equipment	绝缘电阻测试仪 耐压测试仪 电子秒表	DY30-2 WB2670B TF807	0.1MΩ~19GΩ (0~5) kV (0~999.9) s	5级 / U=0.01s k=2	DUOYI /0233684 杭州威博科技 有限公司/083540 深圳市惠波工贸 有限公司/ZT-022
开 展 的 检 定 或 校 准 项 目 Verification or Calibration Items	名 称 Name	测 量 范 围 Measuring Range	不 确 定 度 或 准 确 度 等 级 或 最 大 允 许 误 差 Uncertainty or Accuracy Class or Maximum Permissible Error	依 据 的 计 量 检 定 规 程 或 计 量 技 术 规 范 的 编 号 及 名 称 Verification Regulation or Measurement Technical Specification and Its Code	
	配K型热电偶用 数字温度二次仪表 -----	(0~1300) °C -----	0.2级及以下 -----	JJG 617-1996《数字温度 指示调节仪》 -----	