

计量标准考核证书

Certificate for Examination of Measurement Standard

[2019] 长 量标 企 证字 第 295 号

根据《中华人民共和国计量法》，按照《计量标准考核规范》的要求，考核合格，特发此证。

This is to certify that the measurement standard conforms with the requirements of the "Rule for the Examination of Measurement Standards" according to "the Law on Metrology of the People's Republic of China".

建标单位名称 长沙中通检测技术服务有限公司
Possessor of the Measurement Standard

计量标准名称 耐电压测试仪检定装置
Name of Measurement Standard

代码 15611103
Code

测量范围 直流高压：(2~10) kV，交流高压：(2~10) kV，
Measuring Range 直流击穿电流：(0.1~100) mA，交流击穿电流：(0.1~100) mA，时间：10s~90s

不确定度或准确度等级 直流高压：(2~10) kV MPE: ±(0.5%读数+5字)
或最大允许误差 交流高压：(2~10) kV MPE: ±(1.0%读数+5字)
Uncertainty or Accuracy Class 直流击穿电流：(0.1~100) mA MPE: ±(0.05%读数+2字)
Maximum Permissible Error 交流击穿电流：(0.1~100) mA MPE: ±(0.5%读数+10字)
时 间： 10s~90s MPE: ±0.5s/d

保存地点 电学室
Installed in

发证机关 (印章)
The Issuing Authority

发证日期 2019 年 07 月 31 日
Date of Issue

有效期至 2023 年 07 月 30 日
Date of Expiry

Measurement Standard	名称 Name	型号 Type	测量范围 Measuring Range	不确定度或准确度等级 或最大允许误差 Uncertainty or Accuracy Class or Maximum Permissible Error	制造厂及 出厂编号 Manufacturer and Series Number
	计量 标准 器	电子秒表	TF807	10s~90s	MPE: ±0.5s/d
数字高压表		149-10A	ACV: (2~10) kV; DCV: (2~10) kV	MPE: ±(1.0%读数+5字) MPE: ±(0.5%读数+5字)	KIKUSUI 25100817
数字万用表		45	ACA: (0.1~100) mA; DCA: (0.1~100) mA	MPE: ±(0.5%读数+10字) MPE: ±(0.05%读数+2字)	FLUKE 87311421048
Main Auxiliary Equipment	绝缘电阻	DY30-2	0 Ω~10G Ω	MPE: ±5%rdg	DUOYI 0233684
Verification or Calibration Items	名称 Name	测量范围 Measuring Range	不确定度或准确度等级 或最大允许误差 Uncertainty or Accuracy Class or Maximum Permissible Error	依据的计量检定规程 或计量技术规范的编号及名称 Verification Regulation or Measurement Technical Specification and Its Code	
	耐电压测试仪 (后续检定)	直流高压: (2~10) kV; 交流高压: (2~10) kV; 直流击穿电流: (0.1~100) mA; 交流击穿电流: (0.1~100) mA; 时间: (10~90) s	5.0级及以下	JJG 795-2016 耐电压测试仪检定规程	